

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055923.D
 Acq On : 31 May 2019 16:49
 Operator : JC/SP
 Sample : K3129-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z1-TB02-190528

Quant Time: Jun 01 03:11:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	349285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	532418	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	481621	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	154949	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	190842	52.91	ug/l	0.00
Spiked Amount						
			Recovery	=	105.82%	
35) Dibromofluoromethane	7.59	113	165339	49.12	ug/l	0.00
Spiked Amount						
			Recovery	=	98.24%	
50) Toluene-d8	10.09	98	664177	51.44	ug/l	0.00
Spiked Amount						
			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	12.40	95	206330	46.60	ug/l	0.00
Spiked Amount						
			Recovery	=	93.20%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055923.D
Acq On : 31 May 2019 16:49
Operator : JC/SP
Sample : K3129-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z1-TB02-190528

Quant Time: Jun 01 03:11:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

