

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055925.D
 Acq On : 31 May 2019 17:39
 Operator : JC/SP
 Sample : K3129-11 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF012MW033-190529

Quant Time: Jun 01 03:18:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	358410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	534161	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	483606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	151100	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	193824	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
35) Dibromofluoromethane	7.59	113	170120	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
50) Toluene-d8	10.09	98	681307	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	12.40	95	206852	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
Target Compounds						
4) Vinyl Chloride	2.18	62	810507	224.848	ug/l	99
12) 1,1-Dichloroethene	3.72	96	2186	0.717	ug/l	83
27) cis-1,2-Dichloroethene	6.83	96	1437683	376.302	ug/l	90
65) Chlorobenzene	11.43	112	24062	2.485	ug/l	100
91) 1,2-Dichlorobenzene	13.65	146	5480	1.076	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055925.D
Acq On : 31 May 2019 17:39
Operator : JC/SP
Sample : K3129-11 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF012MW033-190529

Quant Time: Jun 01 03:18:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

