

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055929.D
 Acq On : 31 May 2019 19:19
 Operator : JC/SP
 Sample : K3129-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF012RW160-190529

Quant Time: Jun 01 03:42:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	359177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	537346	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	494401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	163131	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	192191	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
35) Dibromofluoromethane	7.59	113	167982	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	10.09	98	676138	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.40	95	217428	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
Target Compounds						
4) Vinyl Chloride	2.19	62	46365	12.835	ug/l	99
27) cis-1,2-Dichloroethene	6.83	96	330658	86.362	ug/l	90
65) Chlorobenzene	11.43	112	17450	1.763	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055929.D
Acq On : 31 May 2019 19:19
Operator : JC/SP
Sample : K3129-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF012RW160-190529

Quant Time: Jun 01 03:42:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

