

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055942.D
 Acq On : 1 Jun 2019 2:26
 Operator : JC/SP
 Sample : K3125-14 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: Jun 01 04:13:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	350606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	528860	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	494790	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	164967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	177560	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	7.59	113	151884	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
50) Toluene-d8	10.09	98	630215	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	12.40	95	206813	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
Target Compounds						
16) Acetone	3.82	43	348458	218.933	ug/l	99
25) 2-Butanone	6.85	43	5308	2.388	ug/l #	88
52) Toluene	10.16	92	545573	62.452	ug/l	100
67) Ethyl Benzene	11.51	91	16303	0.918	ug/l	98
68) m/p-Xylenes	11.62	106	25717	3.830	ug/l	97
69) o-Xylene	11.95	106	9488	1.424	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055942.D
Acq On : 1 Jun 2019 2:26
Operator : JC/SP
Sample : K3125-14 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: Jun 01 04:13:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

