

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
 Data File : VN077990.D
 Acq On : 31 May 2023 14:08
 Operator : JC\MD
 Sample : 02943-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D2-20230523

Quant Time: Jun 01 01:51:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	392199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	721074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	637622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	199599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	279028	47.548	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.100%
35) Dibromofluoromethane	8.177	113	235784	50.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.300%
50) Toluene-d8	10.571	98	902546	49.869	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.740%
62) 4-Bromofluorobenzene	12.859	95	291187	45.555	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	75865	14.763	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
Data File : VN077990.D
Acq On : 31 May 2023 14:08
Operator : JC\MD
Sample : 02943-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D2-20230523

Quant Time: Jun 01 01:51:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

