

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
 Data File : VN077991.D
 Acq On : 31 May 2023 14:32
 Operator : JC\MD
 Sample : 02973-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-140N-20230526

Quant Time: Jun 01 01:51:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	380249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	697375	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	600181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	176045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	266023	46.757	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.520%
35) Dibromofluoromethane	8.177	113	230062	50.591	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.571	98	852182	48.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.859	95	266607	43.127	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.260%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	4039	2.256	ug/l #	86
27) cis-1,2-Dichloroethene	7.494	96	2713	0.504	ug/l	97
44) Trichloroethene	9.359	130	4221	0.849	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
Data File : VN077991.D
Acq On : 31 May 2023 14:32
Operator : JC\MD
Sample : 02973-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-140N-20230526

Quant Time: Jun 01 01:51:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

