

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
 Data File : VN077997.D
 Acq On : 31 May 2023 16:55
 Operator : JC\MD
 Sample : 02973-11DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INTERCEPTOR-SUMP-20230526DL

Quant Time: Jun 01 01:55:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	379556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	696071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	582643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	164162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257475	45.337	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.680%		
35) Dibromofluoromethane	8.177	113	222135	48.939	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.880%		
50) Toluene-d8	10.576	98	846038	48.425	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.860%		
62) 4-Bromofluorobenzene	12.859	95	258135	41.835	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.660%		
Target Compounds						
27) cis-1,2-Dichloroethene	7.494	96	57967	10.781	ug/l	97
44) Trichloroethene	9.359	130	329850	66.492	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
Data File : VN077997.D
Acq On : 31 May 2023 16:55
Operator : JC\MD
Sample : 02973-11DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INTERCEPTOR-SUMP-20230526DL

Quant Time: Jun 01 01:55:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

