

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
 Data File : VN078002.D
 Acq On : 31 May 2023 18:55
 Operator : JC\MD
 Sample : 02973-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPRING-20230526

Quant Time: Jun 01 01:57:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	370752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	675784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	578109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	169410	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	254159	45.816	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.640%
35) Dibromofluoromethane	8.177	113	221071	50.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.340%
50) Toluene-d8	10.576	98	824142	48.588	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.859	95	257331	42.956	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
Data File : VN078002.D
Acq On : 31 May 2023 18:55
Operator : JC\MD
Sample : 02973-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPRING-20230526

Quant Time: Jun 01 01:57:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

