

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
 Data File : VN078003.D
 Acq On : 31 May 2023 19:19
 Operator : JC\MD
 Sample : 02943-04 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D2-20230523

Quant Time: Jun 01 01:58:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	360933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	669203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	600248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	191185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	256359	47.470	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.940%
35) Dibromofluoromethane	8.177	113	219956	50.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	10.576	98	838339	49.911	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.820%
62) 4-Bromofluorobenzene	12.859	95	273420	46.091	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.180%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.383	101	6551	1.640	ug/l #	89
44) Trichloroethene	9.359	130	360326	75.552	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053123\
Data File : VN078003.D
Acq On : 31 May 2023 19:19
Operator : JC\MD
Sample : 02943-04 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D2-20230523

Quant Time: Jun 01 01:58:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

