

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082398.D
 Acq On : 31 May 2024 11:00
 Operator : JC\MD
 Sample : VN0531MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0531MBL01

Quant Time: May 31 22:53:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	279337	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	263930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110474	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	128708	51.649	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.300%
35) Dibromofluoromethane	8.165	113	94378	54.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.840%
50) Toluene-d8	10.565	98	355669	53.987	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.853	95	117877	50.130	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082398.D
Acq On : 31 May 2024 11:00
Operator : JC\MD
Sample : VN0531MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0531MBL01

Quant Time: May 31 22:53:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

