

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082402.D
 Acq On : 31 May 2024 12:45
 Operator : JC\MD
 Sample : PB161190TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161190TB

Quant Time: May 31 22:56:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155932	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132530	53.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.660%	
35) Dibromofluoromethane	8.171	113	93032	52.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.560%	
50) Toluene-d8	10.565	98	373192	55.224	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.440%	
62) 4-Bromofluorobenzene	12.847	95	124036	51.424	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 102.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	74765	74.022	ug/l	90
20) Methylene Chloride	5.277	84	13474	6.612	ug/l #	74
25) 2-Butanone	7.483	43	12224	7.982	ug/l #	84
43) Isopropyl Acetate	8.688	43	244767	41.754	ug/l #	75
51) 4-Methyl-2-Pentanone	10.447	43	19094	5.733	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082402.D
Acq On : 31 May 2024 12:45
Operator : JC\MD
Sample : PB161190TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161190TB

Quant Time: May 31 22:56:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

