

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082404.D
 Acq On : 31 May 2024 13:33
 Operator : JC\MD
 Sample : PB161190ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161190ZHE#02

Quant Time: May 31 22:57:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	312293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	132968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	145240	53.465	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.940%	
35) Dibromofluoromethane	8.171	113	103509	53.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.760%	
50) Toluene-d8	10.565	98	412339	55.984	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.960%	
62) 4-Bromofluorobenzene	12.847	95	143163	54.459	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 108.920%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	72291	64.871	ug/l	93
20) Methylene Chloride	5.283	84	14107	6.274	ug/l #	89
25) 2-Butanone	7.489	43	12074	7.146	ug/l #	76
43) Isopropyl Acetate	8.688	43	263237	41.200	ug/l #	73
51) 4-Methyl-2-Pentanone	10.447	43	19500	5.372	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082404.D
Acq On : 31 May 2024 13:33
Operator : JC\MD
Sample : PB161190ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161190ZHE#02

Quant Time: May 31 22:57:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

