

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082406.D
 Acq On : 31 May 2024 14:22
 Operator : JC\MD
 Sample : PB161190ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161190ZHE#04

Quant Time: May 31 22:57:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	165163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	307488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	301654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	140727	53.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.920%	
35) Dibromofluoromethane	8.171	113	100996	53.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.780%	
50) Toluene-d8	10.571	98	403860	55.690	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.380%	
62) 4-Bromofluorobenzene	12.847	95	139389	53.852	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 107.700%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	73363	68.574	ug/l	96
20) Methylene Chloride	5.283	84	13045	6.044	ug/l #	83
25) 2-Butanone	7.488	43	12867	7.933	ug/l #	74
43) Isopropyl Acetate	8.688	43	313683	49.863	ug/l #	71
51) 4-Methyl-2-Pentanone	10.447	43	18036	5.046	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082406.D
Acq On : 31 May 2024 14:22
Operator : JC\MD
Sample : PB161190ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161190ZHE#04

Quant Time: May 31 22:57:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

