

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082408.D
 Acq On : 31 May 2024 15:10
 Operator : JC\MD
 Sample : PB161190ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161190ZHE#06

Quant Time: May 31 22:58:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	157132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133076	53.635	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.260%	
35) Dibromofluoromethane	8.165	113	95061	53.182	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.360%	
50) Toluene-d8	10.571	98	377762	55.124	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.240%	
62) 4-Bromofluorobenzene	12.847	95	124103	50.737	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.480%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	75762	74.436	ug/l	93
20) Methylene Chloride	5.283	84	12095	5.890	ug/l #	73
25) 2-Butanone	7.488	43	12627	8.183	ug/l #	74
43) Isopropyl Acetate	8.688	43	326972	55.001	ug/l #	71
51) 4-Methyl-2-Pentanone	10.447	43	18209	5.391	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082408.D
Acq On : 31 May 2024 15:10
Operator : JC\MD
Sample : PB161190ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161190ZHE#06

Quant Time: May 31 22:58:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

