

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
 Data File : VN082414.D
 Acq On : 31 May 2024 17:35
 Operator : JC\MD
 Sample : P2646-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPS

Quant Time: May 31 23:00:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134302	53.889	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.780%	
35) Dibromofluoromethane	8.171	113	97231	54.805	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.620%	
50) Toluene-d8	10.571	98	379793	55.837	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.680%	
62) 4-Bromofluorobenzene	12.847	95	128857	53.077	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 106.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	76809	75.130	ug/l	89
20) Methylene Chloride	5.271	84	13827	6.703	ug/l	91
25) 2-Butanone	7.488	43	11749	7.580	ug/l #	83
43) Isopropyl Acetate	8.694	43	267016	45.254	ug/l #	72
51) 4-Methyl-2-Pentanone	10.447	43	17683	5.275	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053124\
Data File : VN082414.D
Acq On : 31 May 2024 17:35
Operator : JC\MD
Sample : P2646-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPS

Quant Time: May 31 23:00:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

