

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060120\
 Data File : VN061596.D
 Acq On : 1 Jun 2020 15:46
 Operator : JC/MD
 Sample : L2800-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022MW357-200527

Quant Time: Jun 02 04:43:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	287187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	505689	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	503981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	234662	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	173815	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	7.55	113	148012	57.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.40%	
50) Toluene-d8	10.06	98	611450	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	12.37	95	232855	53.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.38%	

Target Compounds

						Qvalue
4) Vinyl Chloride	2.16	62	1004844	209.340	ug/l	99
12) 1,1-Dichloroethene	3.70	96	3968	1.352	ug/l	96
21) trans-1,2-Dichloroethene	5.00	96	19294	5.830	ug/l	93
24) 1,1-Dichloroethane	5.79	63	51453	9.596	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	469662	125.329	ug/l	92
40) Benzene	8.00	78	106883	7.738	ug/l	97
42) 1,2-Dichloroethane	8.08	62	19579	4.234	ug/l	97
44) Trichloroethene	8.80	130	15749	3.710	ug/l	94
52) Toluene	10.13	92	4387	0.487	ug/l	95
64) Tetrachloroethene	10.60	164	2372	0.500	ug/l	90
65) Chlorobenzene	11.41	112	2165148	206.679	ug/l	98
87) 1,3-Dichlorobenzene	13.26	146	7713	0.996	ug/l	98
88) 1,4-Dichlorobenzene	13.34	146	35568	4.510	ug/l	92
91) 1,2-Dichlorobenzene	13.63	146	34439	4.655	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060120\
Data File : VN061596.D
Acq On : 1 Jun 2020 15:46
Operator : JC/MD
Sample : L2800-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022MW357-200527

Quant Time: Jun 02 04:43:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

