

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060120\
 Data File : VN061602.D
 Acq On : 1 Jun 2020 18:06
 Operator : JC/MD
 Sample : L2814-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-052920

Quant Time: Jun 02 05:00:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	280703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	503510	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	517092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	234122	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	175429	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
35) Dibromofluoromethane	7.55	113	143643	56.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.48%	
50) Toluene-d8	10.06	98	631914	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
62) 4-Bromofluorobenzene	12.38	95	233840	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
Target Compounds						
16) Acetone	3.78	43	20168	16.079	ug/l	95
20) Methylene Chloride	4.51	84	14233	3.049	ug/l	93
43) Isopropyl Acetate	8.13	43	61178	9.472	ug/l #	90
95) Naphthalene	15.10	128	21459	1.633	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060120\
Data File : VN061602.D
Acq On : 1 Jun 2020 18:06
Operator : JC/MD
Sample : L2814-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-052920

Quant Time: Jun 02 05:00:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

