


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060120\
 Data File : VN061583.D
 Acq On : 1 Jun 2020 9:22
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Title : SW846 8260
 Last Update : Fri May 29 02:14:10 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	8381	PASS
75	95	30	60	48.4	25442	PASS
95	95	100	100	100.0	52520	PASS
96	95	5	9	6.8	3592	PASS
173	174	0.00	2	1.1	438	PASS
174	95	50	100	79.2	41608	PASS
175	174	5	9	8.1	3360	PASS
176	174	95	101	98.7	41072	PASS
177	176	5	9	6.2	2550	PASS