

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
 Data File : VN078009.D
 Acq On : 01 Jun 2023 13:43
 Operator : JC\MD
 Sample : VN0601WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0601WBL01

Quant Time: Jun 02 01:50:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	383595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	724403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	652955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	219545	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	274332	47.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.600%
35) Dibromofluoromethane	8.177	113	238944	50.584	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	10.577	98	901165	49.563	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.859	95	298785	46.529	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
Data File : VN078009.D
Acq On : 01 Jun 2023 13:43
Operator : JC\MD
Sample : VN0601WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0601WBL01

Quant Time: Jun 02 01:50:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

