

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
 Data File : VN078014.D
 Acq On : 01 Jun 2023 16:50
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 02 01:53:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	435405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	798032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	717623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	249604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	290818	44.640	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.280%
35) Dibromofluoromethane	8.177	113	253958	48.802	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	10.576	98	988748	49.363	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.859	95	315774	44.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.280%
Target Compounds						
52) Toluene	10.635	92	19842	1.419	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
Data File : VN078014.D
Acq On : 01 Jun 2023 16:50
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 02 01:53:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

