

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
 Data File : VN078016.D
 Acq On : 01 Jun 2023 17:38
 Operator : JC\MD
 Sample : 02958-01DL 40X
 Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ORA-0588DL

Quant Time: Jun 02 01:54:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	429521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	782878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.877	117	704576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	238141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	295293	45.948	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.900%
35) Dibromofluoromethane	8.177	113	251150	49.196	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	10.577	98	962875	49.002	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.859	95	328158	47.286	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.580%
Target Compounds						
					Qvalue	
52) Toluene	10.641	92	308039	22.456	ug/l	98
68) m/p-Xylenes	12.077	106	16256	1.567	ug/l	95
84) 1,2,4-Trimethylbenzene	13.488	105	32202	1.898	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060123\
Data File : VN078016.D
Acq On : 01 Jun 2023 17:38
Operator : JC\MD
Sample : 02958-01DL 40X
Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ORA-0588DL

Quant Time: Jun 02 01:54:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

