

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
 Data File : VN061613.D
 Acq On : 2 Jun 2020 12:14
 Operator : JC/MD
 Sample : L2795-01
 Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 217

Quant Time: Jun 03 07:08:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	276787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.54	114	464069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	433119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	235115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	151809	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
35) Dibromofluoromethane	7.54	113	108475	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	10.06	98	541137	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	12.38	95	215798	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
Target Compounds						
					Qvalue	
10) Methyl Iodide	3.88	142	3922	1.067	ug/l	94
16) Acetone	3.78	43	29494	26.919	ug/l	98
18) Methyl Acetate	4.27	43	10951	3.668	ug/l	100
52) Toluene	10.12	92	162291	19.616	ug/l	99
84) 1,2,4-Trimethylbenzene	13.01	105	54441	3.821	ug/l	99
89) n-Butylbenzene	13.59	91	26260	2.085	ug/l #	77
95) Naphthalene	15.10	128	66513	5.041	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
Data File : VN061613.D
Acq On : 2 Jun 2020 12:14
Operator : JC/MD
Sample : L2795-01
Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
217

Quant Time: Jun 03 07:08:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

