

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
 Data File : VN061618.D
 Acq On : 2 Jun 2020 14:11
 Operator : JC/MD
 Sample : L2800-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS032MW061-200527

Quant Time: Jun 03 03:05:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	311729	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	533932	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	523292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	239515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	184287	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
35) Dibromofluoromethane	7.55	113	151586	55.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.94%	
50) Toluene-d8	10.06	98	649194	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	12.38	95	244962	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
Target Compounds						
65) Chlorobenzene	11.41	112	4035	0.371	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060220\
Data File : VN061618.D
Acq On : 2 Jun 2020 14:11
Operator : JC/MD
Sample : L2800-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS032MW061-200527

Quant Time: Jun 03 03:05:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

