

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
 Data File : VN061620.D
 Acq On : 2 Jun 2020 14:58
 Operator : JC/MD
 Sample : L2800-02DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022MW357-200527DL

Manual Integrations
 APPROVED

MMDadoda
 6/3/2020 11:32:26 AM

Quant Time: Jun 03 07:41:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	365943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	608562	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	594154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	282298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	193828	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
35) Dibromofluoromethane	7.55	113	168223	54.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.98%	
50) Toluene-d8	10.06	98	713080	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.37	95	277114	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.16	62	291676	47.688	ug/l	99
12) 1,1-Dichloroethene	3.70	96	1537	0.411	ug/l #	66
21) trans-1,2-Dichloroethene	4.99	96	6714	1.592	ug/l	94
24) 1,1-Dichloroethane	5.80	63	15016	2.198	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	153822	32.213	ug/l	89
40) Benzene	8.00	78	31447	1.892	ug/l	99
42) 1,2-Dichloroethane	8.08	62	6302	1.132	ug/l	88
44) Trichloroethene	8.80	130	5362	1.050	ug/l	93
65) Chlorobenzene	11.41	112	581812	47.109	ug/l	98
88) 1,4-Dichlorobenzene	13.34	146	11396m	1.201	ug/l	
91) 1,2-Dichlorobenzene	13.63	146	9669	1.086	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
Data File : VN061620.D
Acq On : 2 Jun 2020 14:58
Operator : JC/MD
Sample : L2800-02DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022MW357-200527DL

Manual Integrations
APPROVED

MMDadoda
6/3/2020 11:32:26 AM

Quant Time: Jun 03 07:41:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

