

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060220\
 Data File : VN061623.D
 Acq On : 2 Jun 2020 16:49
 Operator : JC/MD
 Sample : L2504-04
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-SLUDGE

Quant Time: Jun 03 07:53:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	308827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.54	114	520273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	508995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	267732	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	173680	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
35) Dibromofluoromethane	7.54	113	129885	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
50) Toluene-d8	10.06	98	625822	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.38	95	247770	55.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.06%	
Target Compounds						
65) Chlorobenzene	11.41	112	31521	2.979	ug/l	98
88) 1,4-Dichlorobenzene	13.34	146	19543	2.172	ug/l	82
91) 1,2-Dichlorobenzene	13.63	146	14969	1.773	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN060220\
Data File : VN061623.D
Acq On : 2 Jun 2020 16:49
Operator : JC/MD
Sample : L2504-04
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-SLUDGE

Quant Time: Jun 03 07:53:01 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

