

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
 Data File : VN078023.D
 Acq On : 02 Jun 2023 11:45
 Operator : JC\MD
 Sample : PB153165TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153165TB

Quant Time: Jun 05 01:16:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	394896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	733894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	665443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	202726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	272685	46.150	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.300%
35) Dibromofluoromethane	8.177	113	235386	49.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	10.571	98	911958	49.508	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.853	95	298105	45.822	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.640%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	44618	24.002	ug/l	95
20) Methylene Chloride	5.295	84	23738	3.051	ug/l	94
37) Ethyl Acetate	7.571	43	76989	12.340	ug/l	97
43) Isopropyl Acetate	8.700	43	494805	48.555	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
Data File : VN078023.D
Acq On : 02 Jun 2023 11:45
Operator : JC\MD
Sample : PB153165TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153165TB

Quant Time: Jun 05 01:16:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

