

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
 Data File : VN078026.D
 Acq On : 02 Jun 2023 12:57
 Operator : JC\MD
 Sample : PB153165ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153165ZHE#03

Quant Time: Jun 05 01:17:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	370093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	703680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	633875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	193059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	259183	46.805	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.600%
35) Dibromofluoromethane	8.177	113	226721	49.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.820%
50) Toluene-d8	10.577	98	869427	49.226	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.460%
62) 4-Bromofluorobenzene	12.853	95	255120	40.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.800%#
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	42650	24.481	ug/l	97
20) Methylene Chloride	5.295	84	23226	3.276	ug/l #	93
37) Ethyl Acetate	7.577	43	77876	13.018	ug/l	97
43) Isopropyl Acetate	8.694	43	655011	67.035	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
Data File : VN078026.D
Acq On : 02 Jun 2023 12:57
Operator : JC\MD
Sample : PB153165ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153165ZHE#03

Quant Time: Jun 05 01:17:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

