

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
 Data File : VN078028.D
 Acq On : 02 Jun 2023 13:45
 Operator : JC\MD
 Sample : 03008-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-18-0A

Quant Time: Jun 05 01:19:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	372525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	705381	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	650343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	210594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	259454	46.548	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.100%
35) Dibromofluoromethane	8.177	113	227120	49.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	10.576	98	883772	49.918	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.859	95	295814	47.308	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.620%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	42783	24.397	ug/l	90
20) Methylene Chloride	5.283	84	18831	2.238	ug/l #	80
37) Ethyl Acetate	7.571	43	70449	11.748	ug/l	98
43) Isopropyl Acetate	8.700	43	810591	82.758	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
Data File : VN078028.D
Acq On : 02 Jun 2023 13:45
Operator : JC\MD
Sample : 03008-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-18-0A

Quant Time: Jun 05 01:19:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

