

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
 Data File : VN078030.D
 Acq On : 02 Jun 2023 14:33
 Operator : JC\MD
 Sample : 03001-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP16-0A

Quant Time: Jun 05 01:19:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	367810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	690164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	630355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	198568	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	254741	46.288	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.580%		
35) Dibromofluoromethane	8.177	113	217421	48.311	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.620%		
50) Toluene-d8	10.576	98	858263	49.546	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.100%		
62) 4-Bromofluorobenzene	12.859	95	288146	47.098	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.200%		
Target Compounds						Qvalue
16) Acetone	4.453	43	72440	41.838	ug/l	96
20) Methylene Chloride	5.288	84	27732	4.350	ug/l	99
25) 2-Butanone	7.500	43	13827	5.105	ug/l	90
37) Ethyl Acetate	7.571	43	58512	9.973	ug/l	97
40) Benzene	8.612	78	253892	13.070	ug/l	98
43) Isopropyl Acetate	8.700	43	630485m	65.789	ug/l	
73) Isopropylbenzene	12.706	105	17281	1.010	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
Data File : VN078030.D
Acq On : 02 Jun 2023 14:33
Operator : JC\MD
Sample : 03001-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP16-0A

Quant Time: Jun 05 01:19:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

