

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
 Data File : VN078031.D
 Acq On : 02 Jun 2023 14:57
 Operator : JC\MD
 Sample : 03001-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP17-0A

Quant Time: Jun 05 01:20:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	413011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	782704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	717413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	236926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	287915	46.590	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.180%		
35) Dibromofluoromethane	8.177	113	246963	48.387	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.780%		
50) Toluene-d8	10.576	98	971603	49.457	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.920%		
62) 4-Bromofluorobenzene	12.853	95	325943	46.977	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.960%		
Target Compounds						Qvalue
16) Acetone	4.447	43	63252	32.533	ug/l	96
20) Methylene Chloride	5.283	84	25461	3.181	ug/l #	82
37) Ethyl Acetate	7.571	43	70862	10.650	ug/l	99
40) Benzene	8.618	78	122825	5.575	ug/l	97
43) Isopropyl Acetate	8.700	43	557864	51.329	ug/l #	69
95) Naphthalene	15.647	128	41796	2.581	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060223\
Data File : VN078031.D
Acq On : 02 Jun 2023 14:57
Operator : JC\MD
Sample : 03001-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP17-0A

Quant Time: Jun 05 01:20:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

