

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060319\
 Data File : VN055954.D
 Acq On : 3 Jun 2019 12:09
 Operator : JC/SP
 Sample : K3129-11DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF012MW033-190529DL

Quant Time: Jun 03 17:53:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	352617	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	535421	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	501149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	163266	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	189654	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
35) Dibromofluoromethane	7.59	113	166814	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	10.09	98	682789	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
62) 4-Bromofluorobenzene	12.40	95	215080	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
Target Compounds						
4) Vinyl Chloride	2.18	62	58072	16.375	ug/l	99
27) cis-1,2-Dichloroethene	6.82	96	127338	33.877	ug/l	91
65) Chlorobenzene	11.43	112	2305	0.230	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060319\
Data File : VN055954.D
Acq On : 3 Jun 2019 12:09
Operator : JC/SP
Sample : K3129-11DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF012MW033-190529DL

Quant Time: Jun 03 17:53:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

