

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060319\
 Data File : VN055960.D
 Acq On : 3 Jun 2019 14:40
 Operator : JC/SP
 Sample : K3125-16 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-3

Quant Time: Jun 03 19:34:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	357257	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	537323	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	495545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	163237	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	179047	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	7.59	113	158652	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
50) Toluene-d8	10.09	98	644561	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	12.40	95	210927	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
Target Compounds						
16) Acetone	3.82	43	357220	220.260	ug/l #	82
52) Toluene	10.16	92	90455	10.191	ug/l	100
67) Ethyl Benzene	11.51	91	3763	0.212	ug/l	92
68) m/p-Xylenes	11.62	106	6992	1.040	ug/l	98
69) o-Xylene	11.95	106	2870	0.430	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060319\
Data File : VN055960.D
Acq On : 3 Jun 2019 14:40
Operator : JC/SP
Sample : K3125-16 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-3

Quant Time: Jun 03 19:34:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

