

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060319\
 Data File : VN055963.D
 Acq On : 3 Jun 2019 15:55
 Operator : JC/SP
 Sample : K3125-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-2

Quant Time: Jun 04 02:03:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	360655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	538041	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	507954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	188298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	176183	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
35) Dibromofluoromethane	7.59	113	155149	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
50) Toluene-d8	10.09	98	646031	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	12.40	95	217809	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
Target Compounds						
16) Acetone	3.82	43	26908	16.435	ug/l	99
25) 2-Butanone	6.85	43	10428	4.561	ug/l	95
52) Toluene	10.16	92	10969	1.234	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060319\
Data File : VN055963.D
Acq On : 3 Jun 2019 15:55
Operator : JC/SP
Sample : K3125-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-2

Quant Time: Jun 04 02:03:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

