


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060319\
Data File : VN055949.D
Acq On : 3 Jun 2019 8:19
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Title : SW846 8260
Last Update : Fri May 24 06:41:54 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	5581	PASS
75	95	30	60	47.5	15505	PASS
95	95	100	100	100.0	32629	PASS
96	95	5	9	6.8	2234	PASS
173	174	0.00	2	0.9	273	PASS
174	95	50	100	93.2	30421	PASS
175	174	5	9	7.4	2262	PASS
176	174	95	101	98.6	29997	PASS
177	176	5	9	7.3	2182	PASS