

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060320\
 Data File : VN061635.D
 Acq On : 3 Jun 2020 13:59
 Operator : JC/MD
 Sample : PB129263TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129263TB

Quant Time: Jun 04 08:12:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	275921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	468612	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	444299	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	189401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	154702	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
35) Dibromofluoromethane	7.55	113	101174	42.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.12%	
50) Toluene-d8	10.06	98	552859	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.38	95	199754	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
Target Compounds						
30) Chloroform	7.33	83	6507	1.142	ug/l	94
95) Naphthalene	15.11	128	11776	1.108	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060320\
Data File : VN061635.D
Acq On : 3 Jun 2020 13:59
Operator : JC/MD
Sample : PB129263TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129263TB

Quant Time: Jun 04 08:12:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

