

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060320\
 Data File : VN061639.D
 Acq On : 3 Jun 2020 15:32
 Operator : JC/MD
 Sample : L2837-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP

Quant Time: Jun 04 08:21:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	293005	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	513933	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	530860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	237128	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	175097	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.55	113	135341	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
50) Toluene-d8	10.06	98	637122	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
62) 4-Bromofluorobenzene	12.38	95	236631	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
Target Compounds						
16) Acetone	3.78	43	21749	16.822	ug/l	93
20) Methylene Chloride	4.51	84	23214	5.777	ug/l	90
43) Isopropyl Acetate	8.13	43	62390	9.463	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060320\
Data File : VN061639.D
Acq On : 3 Jun 2020 15:32
Operator : JC/MD
Sample : L2837-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP

Quant Time: Jun 04 08:21:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

