

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060324\
 Data File : VN082426.D
 Acq On : 03 Jun 2024 11:46
 Operator : JC\MD
 Sample : VN0603MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 04 00:24:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	274639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	268241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122753	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128720	53.030	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.060%	
35) Dibromofluoromethane	8.171	113	91433	54.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.240%	
50) Toluene-d8	10.571	98	355066	54.818	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.640%	
62) 4-Bromofluorobenzene	12.853	95	125632	54.342	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 108.680%	

Target Compounds				Qvalue		
3) Chloromethane	2.365	50	1168	0.465	ug/l #	85
5) Bromomethane	2.971	94	1810	1.334	ug/l #	61
10) Methyl Iodide	4.595	142	784	0.432	ug/l #	40
13) Acrolein	4.189	56	578	2.002	ug/l #	63
14) Allyl chloride	5.030	41	1106	0.308	ug/l #	62
15) Acrylonitrile	5.724	53	1888	1.871	ug/l	93
16) Acetone	4.430	43	2271	2.281	ug/l	93
17) Carbon Disulfide	4.718	76	5049	0.931	ug/l #	81
18) Methyl Acetate	5.024	43	560	0.226	ug/l #	47
19) Methyl tert-butyl Ether	5.800	73	1278	0.208	ug/l #	56
20) Methylene Chloride	5.277	84	1273	0.634	ug/l #	86
21) trans-1,2-Dichloroethene	5.795	96	694	0.379	ug/l #	89
23) Vinyl Acetate	6.612	43	7363	1.244	ug/l	94
25) 2-Butanone	7.489	43	1676	1.110	ug/l #	77
27) cis-1,2-Dichloroethene	7.477	96	1144	0.530	ug/l	85
28) Bromochloromethane	7.812	49	783	0.297	ug/l #	68
29) Tetrahydrofuran	7.853	42	863	0.890	ug/l #	41
31) Cyclohexane	8.224	56	4507	1.163	ug/l #	38
36) 1,1-Dichloropropene	8.371	75	995	0.351	ug/l #	56
38) Carbon Tetrachloride	8.224	117	16462	5.731	ug/l #	16
40) Benzene	8.606	78	1901	0.225	ug/l #	86
42) 1,2-Dichloroethane	8.677	62	1616	0.497	ug/l #	73
43) Isopropyl Acetate	8.694	43	20984	3.735	ug/l #	55
44) Trichloroethene	9.353	130	880	0.458	ug/l	84
46) Dibromomethane	9.712	93	656	0.458	ug/l #	82
47) Bromodichloromethane	9.894	83	851	0.282	ug/l #	61
48) Methyl methacrylate	9.677	41	990	0.370	ug/l #	43
49) 1,4-Dioxane	9.694	88	317	8.449	ug/l #	1
51) 4-Methyl-2-Pentanone	10.447	43	3860	1.209	ug/l #	75
52) Toluene	10.635	92	5011	0.988	ug/l	100
53) t-1,3-Dichloropropene	10.841	75	2151	0.688	ug/l	95
54) cis-1,3-Dichloropropene	10.318	75	1720	0.520	ug/l	96
55) 1,1,2-Trichloroethane	11.012	97	578	0.314	ug/l	94
56) Ethyl methacrylate	10.865	69	719	0.230	ug/l #	64
57) 1,3-Dichloropropane	11.165	76	1391	0.404	ug/l #	81
58) 2-Chloroethyl Vinyl ether	10.159	63	2018	1.251	ug/l #	80


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060324\
 Data File : VN082426.D
 Acq On : 03 Jun 2024 11:46
 Operator : JC\MD
 Sample : VN0603MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 04 00:24:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
59) 2-Hexanone	11.200	43	3310	1.416	ug/l	74
60) Dibromochloromethane	11.371	129	745	0.366	ug/l	97
61) 1,2-Dibromoethane	11.471	107	1084	0.573	ug/l	87
64) Tetrachloroethene	11.118	164	1758	0.884	ug/l #	91
65) Chlorobenzene	11.894	112	7298	1.232	ug/l	99
67) Ethyl Benzene	11.965	91	4487	0.423	ug/l #	86
68) m/p-Xylenes	12.082	106	3211	0.826	ug/l	98
69) o-Xylene	12.400	106	1050	0.279	ug/l	81
70) Styrene	12.418	104	5368	0.859	ug/l	98
71) Bromoform	12.588	173	740	1.171	ug/l #	74
73) Isopropylbenzene	12.700	105	3183	0.326	ug/l	98
74) N-amyl acetate	12.494	43	1785	0.333	ug/l #	74
75) 1,1,2,2-Tetrachloroethane	12.935	83	1020	0.363	ug/l #	88
76) 1,2,3-Trichloropropane	12.847	75	74879	28.982	ug/l #	1
77) Bromobenzene	12.982	156	2489	1.071	ug/l	98
78) n-propylbenzene	13.041	91	6331	0.548	ug/l	99
79) 2-Chlorotoluene	13.123	91	3833	0.529	ug/l	96
80) 1,3,5-Trimethylbenzene	13.176	105	3200	0.406	ug/l	91
81) trans-1,4-Dichloro-2-b...	12.847	75	74879	67.787	ug/l #	12
82) 4-Chlorotoluene	13.223	91	6296	0.888	ug/l	99
83) tert-Butylbenzene	13.441	119	3002	0.435	ug/l	98
84) 1,2,4-Trimethylbenzene	13.488	105	3715	0.460	ug/l	96
85) sec-Butylbenzene	13.618	105	5491	0.574	ug/l	91
86) p-Isopropyltoluene	13.729	119	4724	0.608	ug/l	96
87) 1,3-Dichlorobenzene	13.735	146	5680	1.349	ug/l	95
88) 1,4-Dichlorobenzene	13.735	146	5680	1.322	ug/l	95
89) n-Butylbenzene	14.059	91	8337	1.194	ug/l	99
90) Hexachloroethane	14.341	117	817	0.567	ug/l	89
91) 1,2-Dichlorobenzene	14.106	146	4979	1.249	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	442	0.765	ug/l #	13
93) 1,2,4-Trichlorobenzene	15.394	180	6166	2.760	ug/l	94
94) Hexachlorobutadiene	15.500	225	3171	3.235	ug/l	91
95) Naphthalene	15.647	128	15278	2.182	ug/l	98
96) 1,2,3-Trichlorobenzene	15.847	180	6391	2.858	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:168 Resp: 153722
Ion Ratio Lower Upper
168 100
99 62.7 48.2 72.4



#3
Chloromethane
Concen: 0.465 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 50 Resp: 1168
Ion Ratio Lower Upper
50 100
52 23.5 25.5 38.3#









#5
Bromomethane
Concen: 1.334 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.018 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 1810
Ion Ratio Lower Upper
94 100
96 55.5 74.8 112.2#



#10
Methyl Iodide
Concen: 0.432 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Tgt Ion:142 Resp: 784
Ion Ratio Lower Upper
142 100
127 0.0 35.8 53.6#
141 0.0 11.4 17.2#








#13
 Acrolein
 Concen: 2.002 ug/l
 RT: 4.189 min Scan# 31
 Delta R.T. 0.012 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 56 Resp: 578
 Ion Ratio Lower Upper
 56 100
 55 39.8 55.8 83.8#



#14
 Allyl chloride
 Concen: 0.308 ug/l
 RT: 5.030 min Scan# 523
 Delta R.T. 0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46


Tgt Ion: 41 Resp: 1106
 Ion Ratio Lower Upper
 41 100
 39 64.7 65.4 98.0#
 76 0.0 37.3 55.9#








#15
Acrylonitrile
Concen: 1.871 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53	Resp: 1888
Ion Ratio	Lower Upper
53 100	
52 89.8	66.1 99.1
51 34.6	29.4 44.2



#16
Acetone
Concen: 2.281 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 43	Resp: 2271
Ion Ratio	Lower Upper
43 100	
58 27.7	25.2 37.8









#17
Carbon Disulfide
Concen: 0.931 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 5049
Ion Ratio Lower Upper
76 100
78 16.3 7.5 11.3#



#18
Methyl Acetate
Concen: 0.226 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 43 Resp: 560
Ion Ratio Lower Upper
43 100
74 0.0 22.2 33.2#









#19
Methyl tert-butyl Ether
Concen: 0.208 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 1278
Ion Ratio Lower Upper
73 100
57 0.0 16.5 24.7#



#20
Methylene Chloride
Concen: 0.634 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Tgt Ion: 84 Resp: 1273
Ion Ratio Lower Upper
84 100
49 119.9 84.9 127.3
51 49.0 27.0 40.6#
86 58.7 52.3 78.5








#21
trans-1,2-Dichloroethene
Concen: 0.379 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 96 Resp: 694
Ion Ratio Lower Upper
96 100
61 114.6 96.5 144.7
98 46.2 51.3 76.9#



#23
Vinyl Acetate
Concen: 1.244 ug/l
RT: 6.612 min Scan# 792
Delta R.T. 0.012 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 43 Resp: 7363
Ion Ratio Lower Upper
43 100
86 10.6 10.2 15.4









#25
2-Butanone
Concen: 1.110 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 1676
Ion Ratio Lower Upper
43 100
72 18.9 25.3 37.9#



#27
cis-1,2-Dichloroethene
Concen: 0.530 ug/l
RT: 7.477 min Scan# 939
Delta R.T. -0.012 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Tgt Ion: 96 Resp: 1144
Ion Ratio Lower Upper
96 100
61 143.2 0.0 268.0
98 40.6 0.0 126.8








#28
Bromochloromethane
Concen: 0.297 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 49 Resp: 783
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.6
130 53.9 66.7 100.1#




#29
Tetrahydrofuran
Concen: 0.890 ug/l
RT: 7.853 min Scan# 1003
Delta R.T. 0.017 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Tgt Ion: 42 Resp: 863
Ion Ratio Lower Upper
42 100
72 6.7 39.3 58.9#
71 10.1 37.4 56.0#








#31
Cyclohexane
Concen: 1.163 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.035 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 4507
Ion Ratio Lower Upper
56 100
69 132.5 24.4 36.6#
84 118.1 77.0 115.4#



#33
1,2-Dichloroethane-d4
Concen: 53.030 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 65 Resp: 128720
Ion Ratio Lower Upper
65 100
67 50.3 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:114 Resp: 274639

Ion Ratio Lower Upper

114 100

63 22.5 0.0 41.0

88 15.4 0.0 31.4



#35

Dibromofluoromethane

Concen: 54.120 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46




Tgt Ion:113 Resp: 91433


Ion Ratio Lower Upper

113 100

111 103.8 83.4 125.2

192 18.3 14.0 21.0





#36

1,1-Dichloropropene

Concen: 0.351 ug/l

RT: 8.371 min Scan# 1

Delta R.T. 0.000 min


Lab File: VN082426.D

Acq: 03 Jun 2024 11:46




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	75	Resp:	995
Ion	Ratio	Lower	Upper
75	100		
110	0.0	17.8	53.5#
77	17.8	25.0	37.4#



#38

Carbon Tetrachloride

Concen: 5.731 ug/l




RT: 8.224 min Scan# 1066


Delta R.T. -0.141 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Tgt Ion:	117	Resp:	16462
Ion	Ratio	Lower	Upper
117	100		
119	5.2	76.5	114.7#
121	0.0	23.9	35.9#





#40

Benzene

Concen: 0.225 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 78 Resp: 1901

Ion Ratio Lower Upper

78 100

77 31.9 20.0 30.0#



#42

1,2-Dichloroethane

Concen: 0.497 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. 0.006 min

Lab File: VN082426.D




Acq: 03 Jun 2024 11:46


Tgt Ion: 62 Resp: 1616

Ion Ratio Lower Upper

62 100


98 0.0 0.0 19.6








#43
Isopropyl Acetate
Concen: 3.735 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 20984
Ion Ratio Lower Upper
43 100
61 2.0 22.8 34.2#
87 0.0 12.0 18.0#



#44
Trichloroethene
Concen: 0.458 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Tgt Ion:130 Resp: 880
Ion Ratio Lower Upper
130 100
95 82.2 0.0 195.4








#46
Dibromomethane
Concen: 0.458 ug/l
RT: 9.712 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 93 Resp: 656
Ion Ratio Lower Upper
93 100
95 112.2 64.8 97.2#
174 88.9 73.4 110.0



#47
Bromodichloromethane
Concen: 0.282 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Tgt Ion: 83 Resp: 851
Ion Ratio Lower Upper
83 100
85 32.5 51.9 77.9#
127 0.0 7.5 11.3#








#48
Methyl methacrylate
Concen: 0.370 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 41 Resp: 990
Ion Ratio Lower Upper
41 100
69 20.5 77.9 116.9#
39 45.3 50.9 76.3#




#49
1,4-Dioxane
Concen: 8.449 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46



Tgt Ion: 88 Resp: 317
Ion Ratio Lower Upper
88 100
43 191.2 19.0 28.6#
58 75.1 50.6 76.0









#50
Toluene-d8
Concen: 54.818 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 98 Resp: 355066
Ion Ratio Lower Upper
98 100
100 64.2 51.8 77.6



#51
4-Methyl-2-Pentanone
Concen: 1.209 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 43 Resp: 3860
Ion Ratio Lower Upper
43 100
58 25.7 33.2 49.8#





#52

Toluene

Concen: 0.988 ug/l

RT: 10.635 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 92 Resp: 5011

Ion Ratio Lower Upper

92 100

91 171.5 136.7 205.1



#53

t-1,3-Dichloropropene

Concen: 0.688 ug/l

RT: 10.841 min Scan# 1511

Delta R.T. 0.006 min

Lab File: VN082426.D




Acq: 03 Jun 2024 11:46


Tgt Ion: 75 Resp: 2151

Ion Ratio Lower Upper

75 100

77 34.5 25.4 38.0









#54
 cis-1,3-Dichloropropene
 Concen: 0.520 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. 0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 1720
 Ion Ratio Lower Upper
 75 100
 77 32.5 25.0 37.4
 39 46.4 34.8 52.2



#55
 1,1,2-Trichloroethane
 Concen: 0.314 ug/l
 RT: 11.012 min Scan# 1540
 Delta R.T. -0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46

Tgt Ion: 97 Resp: 578
 Ion Ratio Lower Upper
 97 100
 83 76.5 70.4 105.6
 85 55.9 45.6 68.4
 99 64.5 50.9 76.3









#56
Ethyl methacrylate
Concen: 0.230 ug/l
RT: 10.865 min Scan# 1515
Delta R.T. -0.012 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 69 Resp: 719
Ion Ratio Lower Upper
69 100
41 75.8 46.6 70.0#
39 0.0 26.5 39.7#



#57
1,3-Dichloropropane
Concen: 0.404 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 76 Resp: 1391
Ion Ratio Lower Upper
76 100
78 21.9 25.8 38.6#









#58
2-Chloroethyl Vinyl ether
Concen: 1.251 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 2018
Ion Ratio Lower Upper
63 100
106 19.7 24.6 36.8#



#59
2-Hexanone
Concen: 1.416 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 43 Resp: 3310
Ion Ratio Lower Upper
43 100
58 37.3 28.3 84.9





#60

Dibromochloromethane

Concen: 0.366 ug/l

RT: 11.371 min Scan# 1

Delta R.T. 0.012 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:129 Resp: 745

Ion Ratio Lower Upper

129 100

127 79.5 38.3 114.8



#61

1,2-Dibromoethane

Concen: 0.573 ug/l

RT: 11.471 min Scan# 1618

Delta R.T. 0.000 min

Lab File: VN082426.D




Acq: 03 Jun 2024 11:46


Tgt Ion:107 Resp: 1084

Ion Ratio Lower Upper

107 100

109 82.4 75.8 113.6









#62
 4-Bromofluorobenzene
 Concen: 54.342 ug/l
 RT: 12.853 min Scan# 1853
 Delta R.T. 0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:	95	Resp:	125632
Ion	Ratio	Lower	Upper
95	100		
174	77.2	0.0	160.6
176	74.3	0.0	152.8



#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 11.871 min Scan# 1686
 Delta R.T. 0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46


Tgt Ion:	117	Resp:	268241
Ion	Ratio	Lower	Upper
117	100		
82	60.7	44.6	66.8
119	32.6	25.9	38.9








#64
Tetrachloroethene
Concen: 0.884 ug/l
RT: 11.118 min Scan# 1558
Delta R.T. 0.012 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:164 Resp: 1758
Ion Ratio Lower Upper
164 100
166 126.8 97.6 146.4
129 112.9 73.7 110.5#
131 93.1 75.8 113.6



#65
Chlorobenzene
Concen: 1.232 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion:112 Resp: 7298
Ion Ratio Lower Upper
112 100
114 31.7 25.7 38.5









#67
Ethyl Benzene
Concen: 0.423 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 4487
Ion Ratio Lower Upper
91 100
106 23.6 25.3 37.9#



#68
m/p-Xylenes
Concen: 0.826 ug/l
RT: 12.082 min Scan# 1722
Delta R.T. 0.012 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion:106 Resp: 3211
Ion Ratio Lower Upper
106 100
91 196.4 159.4 239.0









#69
o-Xylene
Concen: 0.279 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 1050
Ion Ratio Lower Upper
106 100
91 239.0 104.7 314.1



#70
Styrene
Concen: 0.859 ug/l
RT: 12.418 min Scan# 1779
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion:104 Resp: 5368
Ion Ratio Lower Upper
104 100
78 49.3 38.6 58.0
103 52.4 43.8 65.6





#71

Bromoform

Concen: 1.171 ug/l

RT: 12.588 min Scan# 1808

Delta R.T. 0.012 min


Lab File: VN082426.D

Acq: 03 Jun 2024 11:46




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	173	Resp:	740
Ion Ratio	Lower	Upper	
173	100		
175	31.6	24.6	74.0
254	0.0	0.0	0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Tgt Ion:	152	Resp:	122753
Ion Ratio	Lower	Upper	
152	100		
115	59.6	30.4	91.3
150	157.6	0.0	353.2





#73

Isopropylbenzene

Concen: 0.326 ug/l

RT: 12.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 3183

Ion Ratio Lower Upper

105 100

120 27.9 13.4 40.1



#74

N-ethyl acetate

Concen: 0.333 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Tgt Ion: 43 Resp: 1785




Ion Ratio Lower Upper


43 100

70 27.5 39.0 58.4#

55 19.4 21.3 31.9#

61 12.0 20.5 30.7#









#75
1,1,2,2-Tetrachloroethane
Concen: 0.363 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 1020
Ion Ratio Lower Upper
83 100
131 0.0 5.5 16.4#
85 58.2 32.9 98.6



#76
1,2,3-Trichloropropane
Concen: 28.982 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.147 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 75 Resp: 74879
Ion Ratio Lower Upper
75 100
77 1.2 90.5 271.3#









#77
 Bromobenzene
 Concen: 1.071 ug/l
 RT: 12.982 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:	156	Resp:	2489
Ion	Ratio	Lower	Upper
156	100		
77	198.1	100.2	300.6
158	94.7	49.1	147.3



#78
 n-propylbenzene
 Concen: 0.548 ug/l
 RT: 13.041 min Scan# 1885
 Delta R.T. 0.006 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46

Tgt Ion:	91	Resp:	6331
Ion	Ratio	Lower	Upper
91	100		
120	23.6	11.7	35.0









#79
2-Chlorotoluene
Concen: 0.529 ug/l
RT: 13.123 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 3833
Ion Ratio Lower Upper
91 100
126 32.5 17.5 52.5



#80
1,3,5-Trimethylbenzene
Concen: 0.406 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 105 Resp: 3200
Ion Ratio Lower Upper
105 100
120 43.2 24.8 74.3









#81
 trans-1,4-Dichloro-2-butene
 Concen: 67.787 ug/l
 RT: 12.847 min Scan# 1852
 Delta R.T. 0.112 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 74879
 Ion Ratio Lower Upper
 75 100
 53 0.0 72.9 109.3#
 89 0.0 39.1 58.7#



#82
 4-Chlorotoluene
 Concen: 0.888 ug/l
 RT: 13.223 min Scan# 1916
 Delta R.T. 0.000 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46


Tgt Ion: 91 Resp: 6296
 Ion Ratio Lower Upper
 91 100
 126 34.5 17.4 52.3








#83
tert-Butylbenzene
Concen: 0.435 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:119 Resp: 3002
Ion Ratio Lower Upper
119 100
91 66.6 32.6 97.8
134 25.4 12.4 37.0



#84
1,2,4-Trimethylbenzene
Concen: 0.460 ug/l
RT: 13.488 min Scan# 1961
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion:105 Resp: 3715
Ion Ratio Lower Upper
105 100
120 48.5 22.9 68.7





#85

sec-Butylbenzene

Concen: 0.574 ug/l

RT: 13.618 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 5491

Ion Ratio Lower Upper

105 100

134 16.2 10.2 30.4



#86

p-Isopropyltoluene

Concen: 0.608 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46




Tgt Ion:119 Resp: 4724


Ion Ratio Lower Upper

119 100

134 22.9 13.2 39.5

91 24.1 11.8 35.3









#87
 1,3-Dichlorobenzene
 Concen: 1.349 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146 Resp: 5680
 Ion Ratio Lower Upper
 146 100
 111 43.0 20.8 62.5
 148 59.3 32.1 96.3



#88
 1,4-Dichlorobenzene
 Concen: 1.322 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. -0.076 min
 Lab File: VN082426.D
 Acq: 03 Jun 2024 11:46

Tgt Ion:146 Resp: 5680
 Ion Ratio Lower Upper
 146 100
 111 43.0 19.3 57.9
 148 59.3 31.3 93.9





#89

n-Butylbenzene

Concen: 1.194 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 8337

Ion Ratio Lower Upper

91 100

92 52.4 26.2 78.5

134 25.7 13.2 39.5



#90

Hexachloroethane

Concen: 0.567 ug/l

RT: 14.341 min Scan# 2106

Delta R.T. 0.006 min

Lab File: VN082426.D




Acq: 03 Jun 2024 11:46


Tgt Ion: 117 Resp: 817

Ion Ratio Lower Upper

117 100

201 75.4 33.3 99.8









#91
1,2-Dichlorobenzene
Concen: 1.249 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 4979
Ion Ratio Lower Upper
146 100
111 40.1 21.3 64.0
148 63.6 31.8 95.4



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.765 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion: 75 Resp: 442
Ion Ratio Lower Upper
75 100
155 0.0 42.6 128.0#
157 21.0 53.0 159.0#





#93

1,2,4-Trichlorobenzene

Concen: 2.760 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN082426.D





Acq: 03 Jun 2024 11:46

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	180	Resp:	6166
Ion Ratio	Lower	Upper	
180	100		
182	100.3	46.7	140.0
145	33.1	16.2	48.6



#94

Hexachlorobutadiene

Concen: 3.235 ug/l




RT: 15.500 min Scan# 2303


Delta R.T. 0.000 min

Lab File: VN082426.D

Acq: 03 Jun 2024 11:46

Tgt Ion:	225	Resp:	3171
Ion Ratio	Lower	Upper	
225	100		
223	66.8	31.1	93.2
227	70.0	30.6	91.6









#95
Naphthalene
Concen: 2.182 ug/l
RT: 15.647 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:128	Resp:	15278
Ion Ratio	Lower	Upper
128	100	
127	11.7	10.1 15.1
129	11.8	8.9 13.3



#96
1,2,3-Trichlorobenzene
Concen: 2.858 ug/l
RT: 15.847 min Scan# 2362
Delta R.T. 0.006 min
Lab File: VN082426.D
Acq: 03 Jun 2024 11:46

Tgt Ion:180	Resp:	6391
Ion Ratio	Lower	Upper
180	100	
182	93.2	46.0 137.8
145	33.6	17.1 51.3

