

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060324\
 Data File : VN082429.D
 Acq On : 03 Jun 2024 13:01
 Operator : JC\MD
 Sample : P2657-03DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1-AB-VOC DL

Quant Time: Jun 04 00:26:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	300591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	292466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	140120	52.741	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.480%	
35) Dibromofluoromethane	8.171	113	99335	53.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.440%	
50) Toluene-d8	10.571	98	389859	54.993	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.980%	
62) 4-Bromofluorobenzene	12.847	95	135364	53.496	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 107.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	24136	22.146	ug/l	97
20) Methylene Chloride	5.283	84	3196	1.453	ug/l #	78
40) Benzene	8.612	78	174711	18.892	ug/l	99
43) Isopropyl Acetate	8.694	43	51558	8.384	ug/l #	73
52) Toluene	10.629	92	114912	20.706	ug/l	98
67) Ethyl Benzene	11.965	91	24545	2.124	ug/l	99
68) m/p-Xylenes	12.071	106	66253	15.640	ug/l	92
69) o-Xylene	12.400	106	62106	15.136	ug/l	90
80) 1,3,5-Trimethylbenzene	13.176	105	27901	3.243	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	63457	7.209	ug/l	96
95) Naphthalene	15.641	128	3771508	493.712	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060324\
Data File : VN082429.D
Acq On : 03 Jun 2024 13:01
Operator : JC\MD
Sample : P2657-03DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1-AB-VOC DL

Quant Time: Jun 04 00:26:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

