

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
 Data File : VN086846.D
 Acq On : 03 Jun 2025 16:34
 Operator : JC\MD
 Sample : PB168225ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168225ZHE#07

Quant Time: Jun 04 02:01:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	233699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	455449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	421415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	186046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161256	49.935	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.880%
35) Dibromofluoromethane	8.165	113	142407	52.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.000%
50) Toluene-d8	10.565	98	570216	50.730	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.847	95	206325	51.986	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.980%
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	37966	31.430	ug/l	93
18) Methyl Acetate	5.018	43	29606	8.757	ug/l	99
20) Methylene Chloride	5.271	84	6822	2.140	ug/l	88
37) Ethyl Acetate	7.559	43	22138	5.478	ug/l	98
43) Isopropyl Acetate	8.688	43	17707	2.186	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
Data File : VN086846.D
Acq On : 03 Jun 2025 16:34
Operator : JC\MD
Sample : PB168225ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168225ZHE#07

Quant Time: Jun 04 02:01:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

