

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
 Data File : VN086852.D
 Acq On : 03 Jun 2025 18:59
 Operator : JC\MD
 Sample : PB168225ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168225ZHE#13

Quant Time: Jun 04 02:04:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	225219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	441254	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	414229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	186632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	157247	50.527	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.060%		
35) Dibromofluoromethane	8.159	113	136142	51.807	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.620%		
50) Toluene-d8	10.565	98	556501	51.102	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.200%		
62) 4-Bromofluorobenzene	12.847	95	206238	53.635	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 107.280%		
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	38685	33.231	ug/l	95
18) Methyl Acetate	5.012	43	28943	8.883	ug/l	99
20) Methylene Chloride	5.277	84	7030	2.288	ug/l	90
37) Ethyl Acetate	7.559	43	20183	5.155	ug/l	98
43) Isopropyl Acetate	8.682	43	17003	2.166	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
Data File : VN086852.D
Acq On : 03 Jun 2025 18:59
Operator : JC\MD
Sample : PB168225ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168225ZHE#13

Quant Time: Jun 04 02:04:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

