

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
 Data File : VN086854.D
 Acq On : 03 Jun 2025 19:48
 Operator : JC\MD
 Sample : PB168225ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168225ZHE#15

Quant Time: Jun 04 02:05:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	229498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	447218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	422170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	190775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159835	50.401	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.800%		
35) Dibromofluoromethane	8.165	113	138149	51.870	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.740%		
50) Toluene-d8	10.565	98	568447	51.503	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.000%		
62) 4-Bromofluorobenzene	12.847	95	210740	54.075	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 108.160%		
Target Compounds						
					Qvalue	
16) Acetone	4.412	43	44018	37.107	ug/l	94
18) Methyl Acetate	5.012	43	29988	9.032	ug/l	95
20) Methylene Chloride	5.265	84	8594	2.745	ug/l	96
37) Ethyl Acetate	7.559	43	21412	5.396	ug/l	97
43) Isopropyl Acetate	8.688	43	17043	2.142	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
Data File : VN086854.D
Acq On : 03 Jun 2025 19:48
Operator : JC\MD
Sample : PB168225ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168225ZHE#15

Quant Time: Jun 04 02:05:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

