

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
 Data File : VN086859.D
 Acq On : 03 Jun 2025 21:49
 Operator : JC\MD
 Sample : Q2185-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP01-MHB-WC

Quant Time: Jun 04 02:08:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	221364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	429762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	408049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	182763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156008	51.002	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.000%		
35) Dibromofluoromethane	8.165	113	134643	52.607	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.220%		
50) Toluene-d8	10.565	98	548008	51.668	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.340%		
62) 4-Bromofluorobenzene	12.847	95	202871	54.171	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 108.340%		
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	41784	36.518	ug/l	98
18) Methyl Acetate	5.024	43	4172	1.303	ug/l #	87
20) Methylene Chloride	5.265	84	19077	6.318	ug/l	96
43) Isopropyl Acetate	8.688	43	31732	4.151	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060325\
Data File : VN086859.D
Acq On : 03 Jun 2025 21:49
Operator : JC\MD
Sample : Q2185-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP01-MHB-WC

Quant Time: Jun 04 02:08:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

