


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060419\
Data File : VN055974.D
Acq On : 4 Jun 2019 7:53
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Title : SW846 8260
Last Update : Tue Jun 04 11:02:09 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	4255	PASS
75	95	30	60	48.5	12195	PASS
95	95	100	100	100.0	25130	PASS
96	95	5	9	6.4	1606	PASS
173	174	0.00	2	0.6	137	PASS
174	95	50	100	94.9	23848	PASS
175	174	5	9	8.2	1962	PASS
176	174	95	101	96.3	22976	PASS
177	176	5	9	7.3	1688	PASS