

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060420\
 Data File : VN061651.D
 Acq On : 4 Jun 2020 14:27
 Operator : JC/MD
 Sample : PB129284TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129284TB

Quant Time: Jun 05 08:00:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	313188	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	514379	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	502583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	234872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	167019	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
35) Dibromofluoromethane	7.55	113	134026	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
50) Toluene-d8	10.06	98	615155	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.38	95	227616	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
Target Compounds						
16) Acetone	3.78	43	21730	15.309	ug/l	99
20) Methylene Chloride	4.50	84	13778	2.407	ug/l	90
43) Isopropyl Acetate	8.13	43	71184	10.788	ug/l #	92
95) Naphthalene	15.11	128	14562m	1.105	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060420\
Data File : VN061651.D
Acq On : 4 Jun 2020 14:27
Operator : JC/MD
Sample : PB129284TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129284TB

Quant Time: Jun 05 08:00:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

