

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060420\
 Data File : VN061655.D
 Acq On : 4 Jun 2020 16:01
 Operator : JC/MD
 Sample : L2871-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WARREN-TP

Quant Time: Jun 05 08:11:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	296687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	514400	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	520532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	242322	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	171467	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
35) Dibromofluoromethane	7.55	113	137895	52.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.68%	
50) Toluene-d8	10.06	98	621797	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.38	95	236917	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	
Target Compounds						
16) Acetone	3.78	43	23127	17.985	ug/l	96
20) Methylene Chloride	4.51	84	19071m	4.347	ug/l	
43) Isopropyl Acetate	8.13	43	65150	9.873	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060420\
Data File : VN061655.D
Acq On : 4 Jun 2020 16:01
Operator : JC/MD
Sample : L2871-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WARREN-TP

Quant Time: Jun 05 08:11:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

