

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
 Data File : VN067210.D
 Acq On : 4 Jun 2021 21:13
 Operator : JC/MD
 Sample : M2589-28
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B9(2-4)

Quant Time: Jun 05 05:32:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.094	168	149019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	256184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	266396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	106977	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.445	65	95183	46.858	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.720%
35) Dibromofluoromethane	8.029	113	76380	48.093	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.180%
50) Toluene-d8	10.448	98	345501	53.416	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.840%
62) 4-Bromofluorobenzene	12.736	95	112713	45.245	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.500%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	8153	11.632	ug/l	95
20) Methylene Chloride	5.109	84	6638	3.761	ug/l #	91
43) Isopropyl Acetate	8.563	43	18169	4.189	ug/l #	82
95) Naphthalene	15.515	128	9939	6.329	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
Data File : VN067210.D
Acq On : 4 Jun 2021 21:13
Operator : JC/MD
Sample : M2589-28
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B9(2-4)

Quant Time: Jun 05 05:32:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

