

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
 Data File : VN067211.D
 Acq On : 4 Jun 2021 21:37
 Operator : JC/MD
 Sample : M2589-29
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B9(4-6)

Quant Time: Jun 05 05:32:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.094	168	144023	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	242232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	237267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	97888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	90125	45.907	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.820%		
35) Dibromofluoromethane	8.027	113	72317	48.157	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.320%		
50) Toluene-d8	10.449	98	325049	53.148	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.300%		
62) 4-Bromofluorobenzene	12.736	95	102249	43.409	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	86.820%		
Target Compounds					Qvalue	
16) Acetone	4.307	43	6775	10.001	ug/l	92
20) Methylene Chloride	5.117	84	6948	4.073	ug/l	92
43) Isopropyl Acetate	8.571	43	10356	2.525	ug/l #	84
95) Naphthalene	15.512	128	3751	5.210	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
Data File : VN067211.D
Acq On : 4 Jun 2021 21:37
Operator : JC/MD
Sample : M2589-29
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B9(4-6)

Quant Time: Jun 05 05:32:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

