

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
 Data File : VN067215.D
 Acq On : 4 Jun 2021 23:11
 Operator : JC/MD
 Sample : M2589-33
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B10(4-6)

Quant Time: Jun 05 05:34:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	138345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	230530	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	220816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	102134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	84559	44.839	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.680%
35) Dibromofluoromethane	8.029	113	69759	48.812	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.620%
50) Toluene-d8	10.446	98	297886	51.179	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.360%
62) 4-Bromofluorobenzene	12.736	95	98472	43.928	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.860%
Target Compounds						
					Qvalue	
16) Acetone	4.307	43	17063	26.222	ug/l	100
20) Methylene Chloride	5.111	84	6986	4.263	ug/l	97
43) Isopropyl Acetate	8.568	43	7835	2.008	ug/l #	81
95) Naphthalene	15.512	128	23663	9.174	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
Data File : VN067215.D
Acq On : 4 Jun 2021 23:11
Operator : JC/MD
Sample : M2589-33
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B10(4-6)

Quant Time: Jun 05 05:34:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

