

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
 Data File : VN067218.D
 Acq On : 5 Jun 2021 00:21
 Operator : JC/MD
 Sample : M2589-36
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B11(4-6)

Quant Time: Jun 05 05:35:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	132241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	221887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	220326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	98654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	81897	45.432	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.860%		
35) Dibromofluoromethane	8.027	113	67546	49.104	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.200%		
50) Toluene-d8	10.446	98	291278	51.993	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.980%		
62) 4-Bromofluorobenzene	12.736	95	97270	45.082	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.160%		
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	23936	38.482	ug/l	92
17) Carbon Disulfide	4.551	76	4797m	1.302	ug/l	
20) Methylene Chloride	5.101	84	6690	4.271	ug/l	92
25) 2-Butanone	7.348	43	6876	6.904	ug/l	93
43) Isopropyl Acetate	8.566	43	17091	4.550	ug/l #	84
95) Naphthalene	15.509	128	33140	11.312	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060421\
Data File : VN067218.D
Acq On : 5 Jun 2021 00:21
Operator : JC/MD
Sample : M2589-36
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B11(4-6)

Quant Time: Jun 05 05:35:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

