

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
 Data File : VN056002.D
 Acq On : 5 Jun 2019 12:01
 Operator : JC/SP
 Sample : K2640-02 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30234

Quant Time: Jun 06 01:43:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	379381	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	573715	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	536243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	190669	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	176327	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
35) Dibromofluoromethane	7.59	113	161755	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
50) Toluene-d8	10.09	98	673350	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	12.40	95	222905	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
Target Compounds						
11) Tert butyl alcohol	4.81	59	5026	7.667	ug/l #	75
16) Acetone	3.82	43	101729	56.466	ug/l	97
18) Methyl Acetate	4.33	43	8131	1.740	ug/l	97
52) Toluene	10.16	92	54454	5.841	ug/l	99
95) Naphthalene	15.13	128	6566	4.516	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060519\
Data File : VN056002.D
Acq On : 5 Jun 2019 12:01
Operator : JC/SP
Sample : K2640-02 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30234

Quant Time: Jun 06 01:43:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

